

ILS Laboratories

8222 Vickers St, Suite 106, San Diego, CA 92111
(619) 329-3999 | ils-lab.com

Bac Water - 10mL

Tested for: Amino Apex



Aminoapex.com

COA #: COA-2026-DKWWTM
Lot Number: 10BAC626
Accession #: ACC-2026-17825
Labeled Content: 10mL

Analysis Date: 09/11/2026
Appearance: Good
Sample Matrix: BAC Water
Date Received: 08/28/2026

Method: Blind Test (HPLC), Water Analysis

PASS



Scan to verify
authenticity at ils-lab.com
Access Code: T2T4G6HC

Identity

Bac Water



Identification (HPLC)

Analyte	Specification	Result	Unit	Status
Identity (HPLC-RTM)	Benzyl alcohol	Confirmed	-	PASS
Fentanyl Screen	Immunoassay, 50 ng/mL cutoff	Not Detected	-	PASS



Bac Water 10mL - 10BAC626

Water Analysis

Analyte	Specification	Result	Unit	Status
pH	4.5 - 7.0	6.9	pH	PASS
Benzyl Alcohol	0.8% - 1.0%	.818	%	PASS

Endotoxin Testing (USP <85>)

Test	Specification	Result	Status
Endotoxin (USP <85>)	Report Result	0.204 EU/mL	Reported

About this result: Endotoxin is reported as a quantitative value. Acceptable limits vary by product type and matrix, so no universal pass/fail threshold applies to RUO products. This result is below commonly referenced endotoxin thresholds.

Notes & Methodology

- Date Tested: 09/11/2026. Methods: Identification (HPLC); Water Analysis.
- The sample was confirmed to be Bac Water by HPLC. Identification by chromatographic retention time comparison with a reference standard.
- Endotoxin tested per USP <85> kinetic chromogenic method. Acceptance criteria per client specification.
- Accreditation Scope: ILS Laboratories maintains ISO/IEC 17025 accreditation for activities listed in its published PJLA scope. Peptide testing reported on this certificate is performed for Research Use Only.
- This certificate reports analytical results for the sample(s) submitted to ILS Laboratories for testing. All sample identifications, product names, brand names, lot/batch numbers, and labeled content are provided by the submitting party and have not been independently verified. Results reflect only the analytical testing of the specific sample(s) provided and do not constitute an evaluation or endorsement of the manufacturer's processes, quality systems, or overall production. ILS Laboratories makes no representations regarding the intended use, safety, efficacy, regulatory status, or intellectual property ownership of any materials tested.



Issued by ILS Laboratories

(619) 329-3999

ils-lab.com



Dr. Greg Kalyuzhny

Lab Director

09/11/2026

COA #: COA-2026-DKWWTM

Access Code: T2T4G6HC

Verify: portal.ils-lab.com/verify/Z5f4H0oyy1LjGGQL

Issued: 09/11/2026